

## CHATILLON® LG Series Mechanical Force Gauge

### Operating Instructions

The LG Series mechanical force gauge is ideal for a wide range of force testing applications. Based on our popular DPP Series force gauge, the LG Series features a larger, 2.25-inch dial for improved resolution and readability, peak hold button, ergonomic shape and improved accuracy of  $\pm 0.5\%$  of full scale. The LG gauge is an excellent, economical solution for tensile or compression testing.

The durable, plastic housing fits naturally in the hand for a comfortable and firm grip even at maximum loads. Gauges may be mounted to a test stand for even greater control and consistent results in repetitive testing applications. Easy to read concentric dial measures clockwise direction only. Dial rotates 360-degrees for taring. A peak hold button captures peak readings. Available in lbf, kgf or N units of measure. LG Series gauges come complete with protective carrying case and a set of stainless steel attachments. A Certificate of Calibration with Data is supplied standard.



**Chatillon**

A registered trademark of AMETEK inc.

---

### CAUTION

The LG Series is a unidirectional force gauge. When applying a load to the instrument the dial indicator MUST rotate in the clockwise direction.

### OPERATION

The zero point of the gauge may not always be in the 12 o'clock position due to gravity effects and the weight of fixtures or grips on the force gauge shaft. To zero the gauge, rotate the bezel on the dial indicator so that the indicator pointer lines up with the zero on the dial face. On the lighter model force gauges (5 LBF, 2KGF, 20N or less), gently tap the gauge to overcome static friction when taking readings.

Gauges are designed for axial loading. Ensure the load is not applied at an angle or an error can be introduced in the force readings. The gauges are designed to have a 10% tare for fixtures and grips. If this 10% tare is exceeded, the force gauge may not measure up to full scale.

#### ***Compression Testing***

For compression testing, assemble the appropriate attachment or grip to the bottom shaft of the force gauge. For compression testing the shaft should be pushed into the gauge, causing the indicator to rotate in the clockwise direction. Attachments should be applied "finger-tight". Do NOT over-tighten.



#### ***Tensile Testing***

For tension testing, assemble the appropriate hook, attachment or grip to the top shaft of the force gauge. For tension testing the shaft should be pulled out of the top of the force gauge, causing the indicator to rotate in the clockwise direction. Attachments should be applied "finger-tight". Do NOT over-tighten.



---

### CAUTION

Do not apply a load to the LG Series force gauge so the indicator rotates in the counterclockwise direction. If excessive loads are applied in the counterclockwise direction, damage may occur to the force gauge.

### PEAK HOLD

The LG Series force gauge has a maximum hold feature. This allows the peak force to be held after the load is removed.

#### *Activate Peak Hold*

To activate this feature, rotate the knob to "MAX ON" position.

#### *Reset Peak Hold*

To reset the reading, simply rotate the switch to "MAX OFF" and then return back to "MAX ON" position. To deactivate this feature, rotate the knob to the "MAX OFF" position.



### SPECIFICATIONS

Accuracy:	±0.5% of Full Scale
Overload Capacity:	110% of Rated Capacity
Maximum Overload:	150% of Rated Capacity
Tare Capacity:	10% of Rated Capacity
Dial Resolution:	0.5% of Full Scale
Deflection:	10mm (.394 inches) at Full Scale
Operating Temperature:	+40°F to +110°F (+5°C to +45°C)
Instrument Weight:	1.5 lbs (0.67 kg)
Shipping Weight with case:	3.5 lbs (1.59 kg)
Warranty:	2 Years


---

### **TEST STAND MOUNTING**

The LG Series force gauge has two #10-32 female threads for mounting to CHATILLON test stands. Ensure the #10-32 screws are not too long, otherwise it may bind and damage the force gauge.

The screw thread length going into the back of the gauge should not exceed 3/8 inch.

The force gauge also has four M3 x 0.5 threaded holes for alternative mounting.



## ORDERING INFORMATION

### Decimal Dials (LBF)

Model	Capacity x Resolution
LG-002	2 x 0.01 lbf
LG-005	5 x 0.025 lbf
LG-010	10 x 0.05 lbf
LG-020	20 x 0.1 lbf
LG-030	30 x 0.15 lbf
LG-050	50 x 0.25 lbf
LG-075	75 x 0.375 lbf
LG-100	100 x 0.5 lbf

### S.I. Dials (Newtons)

Model	Capacity x Resolution
LG-10N	10 x 0.05 N
LG-20N	20 x 0.1 N
LG-30N	30 x 0.15 N
LG-50N	50 x 0.25 N
LG-100N	100 x 0.5 N
LG-200N	200 x 1 N
LG-300N	300 x 1.5 N
LG-500N	500 x 2.5 N

### Metric Dials (KGF)

Model	Capacity x Resolution
LG-1KG	1 x 0.005 kgf
LG-2KG	2 x 0.01 kgf
LG-3KG	3 x 0.015 kgf
LG-5KG	5 x 0.025 kgf
LG-10KG	10 x 0.05 kgf
LG-20KG	20 x 0.1 kgf
LG-30KG	30 x 0.15 kgf
LG-50KG	50 x 0.25 kgf

## ACCESSORIES

Part No.	Description
SPK-FMG-008A	Chisel Point
SPK-FMG-009A	Point
SPK-FMG-010A	Notch
SPK-FMG-011A	Flat
SPK-FMG-012A <sup>1</sup>	Hook, Stationary, 50 lbf
SPK-FMG-012B <sup>1</sup>	Hook, Stationary, 100 lbf
SPK-FMG-013A	Extension Rod, 6-inch
NC002710	Carrying Case
NC002711	Operator's Manual
ML3867 <sup>2</sup>	Hook, Swivel, 20 lbf
ML3850 <sup>2</sup>	Hook, Swivel, 100 lbf
NC000988-3 <sup>2</sup>	Extension Rod, 3"

Notes: <sup>1</sup> Capacity dependent.

<sup>2</sup> Optional accessories.

## MAINTENANCE

Gauge accuracy should be checked periodically to ensure the gauge is within its calibration limits. The calibration can be verified by applying known weights (adjusted for local gravity) to the extension hook. If adjustment is required, the gauge should be returned to AMETEK for calibration.

---

## SPARE PARTS

Description	Part Number
Dial Indicator Assembly, LG-002	NC002677-01
Dial Indicator Assembly, LG-005	NC002677-02
Dial Indicator Assembly, LG-010	NC002677-03
Dial Indicator Assembly, LG-020	NC002677-04
Dial Indicator Assembly, LG-030	NC002677-05
Dial Indicator Assembly, LG-050	NC002677-06
Dial Indicator Assembly, LG-075	NC002677-07
Dial Indicator Assembly, LG-100	NC002677-08

Description	Part Number
Dial Indicator Assembly, LG-1KG	NC002677-13
Dial Indicator Assembly, LG-2KG	NC002677-14
Dial Indicator Assembly, LG-3KG	NC002677-15
Dial Indicator Assembly, LG-5KG	NC002677-16
Dial Indicator Assembly, LG-10KG	NC002677-17
Dial Indicator Assembly, LG-20KG	NC002677-18
Dial Indicator Assembly, LG-30KG	NC002677-19
Dial Indicator Assembly, LG-50KG	NC002677-20

Description	Part Number
Dial Indicator Assembly, LG-10N	NC002677-24
Dial Indicator Assembly, LG-20N	NC002677-25
Dial Indicator Assembly, LG-30N	NC002677-26
Dial Indicator Assembly, LG-50N	NC002677-27
Dial Indicator Assembly, LG-100N	NC002677-28
Dial Indicator Assembly, LG-200N	NC002677-29
Dial Indicator Assembly, LG-300N	NC002677-30
Dial Indicator Assembly, LG-500N	NC002677-31

Description	Part Number
Bezel, Dial Indicator	NC002716
Bearing Support Assembly	NC002715
Top Case (Housing)	NC002713
Bottom Case (Housing)	NC002714
Maximum Peak Hold Knob Assembly	NC002678
Gauge Mounting Plate	NC002671
Label, Gauge Model	NC002621
Ball Bushing	2630
Retaining Ring, Ball Bushing	13007

---

### **PRODUCT WARRANTY**

This instrument is warranted against defects in workmanship, material and design for two (2) years from the date of delivery to the extent that AMETEK will, at its sole option, repair or replace the instrument or any part thereof which is defective provided, however, that the warranty shall not apply to instruments subjected to tampering or abuse, or exposed to highly corrosive conditions.

THIS WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES WHETHER EXPRESS OR IMPLIED AND AMETEK HEREBY DISCLAIMS ALL OTHER WARRANTIES, INCLUDING, WITHOUT LIMITATION, ANY WARRANTY OF FITNESS FOR A PARTICULAR PURPOSE OR MERCHANTABILITY. AMETEK SHALL NOT BE LIABLE FOR ANY INCIDENTAL OR CONSEQUENTIAL DAMAGES, INCLUDING, BUT NOT LIMITED TO, ANY ANTICIPATED OR LOST PROFITS.

This warranty is voidable if the purchaser fails to follow any and all instructions, warnings or cautions in the instrument's operation manual.

If a manufacturing defect is found, AMETEK will replace or repair the instrument or replace any defective part thereof without charge; however, AMETEK's obligation hereunder does not include the cost of transportation which must be borne by the customer. AMETEK assumes no responsibility for damage in transit, and any claims for such damage should be presented to the carrier by the purchaser.

---

# **AMETEK**

## **TEST AND CALIBRATION INSTRUMENTS**

---

### **Americas**

#### **AMETEK Test and Calibration Instruments Division**

8600 Somerset Drive  
Largo, Florida 33773  
USA

Tel +1-727-536-7831  
Tel +1-800-527-9999 (USA only)  
Fax +1-727-539-6882

### **Asia Pacific**

#### **AMETEK Lloyd Instruments**

No. 7 Sherwood Place  
Alexander Heights  
6064 Perth  
Australia

Tel +61-8-9343-5725  
Fax +61-8-9343-5723

#### **AMETEK Singapore Pvt. Ltd.**

10 Ang Mo Kio Street 65  
#05-12 TECHPOINT  
Singapore  
569059  
Tel +65 484 2388  
Fax +65 481 6588

AMETEK and CHATILLON are registered  
trademarks of AMETEK, Inc.  
LLOYD INSTRUMENTS is an AMETEK Inc.  
trademark  
*Information within this document is  
subject to change without notice.*

Pub No. NC002711  
Issued 01/03

### **Europe**

#### **AMETEK Lloyd Instruments**

Forum House  
12 Barnes Wallis Road  
Segensworth East  
Fareham

Hampshire PO 15 5TT  
Tel +44 (0) 1489-486399  
Fax +44 (0) 1489-8851118

#### **AMETEK Lloyd Instruments SA**

3 Avenue des Coudriers  
Zone d'Activite de l'Observatoire  
7810 Montigny-Le-Bretonneux  
France  
Tel +33-1-3057-4774  
Fax +33-1-3057-5033

#### **AMETEK Precision**

#### **Instruments Europe GmbH**

Rudolf-Diesel-Strasse 16  
D-40670, Meerbusch  
Germany  
Tel +49-2-159-9136 0  
Fax +49-2-159-9136 39

#### **Internet Addresses:**

[www.ametek.com](http://www.ametek.com)  
[www.chatillon.com](http://www.chatillon.com)  
[www.lloyd-instruments.co.uk](http://www.lloyd-instruments.co.uk)

**ISO 9001  
Manufacturer**

Copyright 2003, by AMETEK, Inc.

Printed in U.S.A.



